

Figure S1. (**A**) Technical variability for f-LAESI-MS based on sampling 1 μ L droplets (n = 15) of 500 μ M glutamate standard solution. Results are consistent with a normal distribution and a COV of 14.9%. Single cell results are a convolution of technical and biological variability. For example, sampling (**B**) *E. densa* epidermal cells (n = 97) and (**C**) *G. max* infected root nodule cells (n = 60) revealed gamma distributions with COV values of 97.3% and 50.7%, respectively.