

Figure S3. The ct-HSQC (a) has reduced sensitivity compared to the standard HSQC experiment (b): as illustrated here, the peak at 6.06/88.9 ppm is not visible in the ct-HSQC spectrum.

